In the USA, Federal Law restricts this device to sale by, or on the order of a physician or other practitioner licensed to use the device. For use with a neurostimulator approved for distribution within the United States.

MAINTENANCE:

Check the Conductive Garment before each use for signs of wear and/or damage. Replace StimGel Pads as required. See below for StimGel maintenance suggestions. See label on garment for washing instructions.

WARNINGS AND PRECAUTIONS:

For all relevant hazards, contraindications, warnings, precautions and adverse side effects, refer to the Operators Manual of the electrotherapy device. Using StimGel Pads that are too small for your device could result in thermal burns. Garment should never be used without StimGel pads. The back wrap is intended for single-patient use only. Do not share the wrap with other persons because of the risks of adverse skin reactions and disease transmission.

CAUTION:

Do not apply to broken skin. Should a skin rash or skin burn occur, immediately discontinue use and contact your physician.

90 DAY WARRANTY:

Empi will repair or replace, at its facility, any product found to be defective. This warranty does not apply to any product damaged by misuse, or repaired or altered by anyone other than Empi in Clear Lake, SD.

This warranty is in lieu of any or all other warranties expressed or implied. No person is authorized to bind Empi to any representative of warranty other than those specifically set forth herein.

NOTE: Warranty period begins with the date of purchase from manufacturer.

Specialized Back Wrap

Part Number 235666

Empi

INSTRUCTIONS FOR USE

Please read your device user's manual before use.

Empi

Empi, Inc.
205 Highway 22 East
Clear Lake, SD 57226

www.empi.com
800.328.2536

Made in Mexico for Empi.

360358 Rev C ©2011 Empi, 06/13 All Rights Reserved.
Caution: Federal law restricts this device to sale by or on the order of a physician.

Latex Free

DJO GLOBAL

To reorder call Empi at **800.328.2536**

Empi 99599-001 (pack of four)

Para Español: 360421
Pour le Français: 360420

INSTRUCTIONS FOR USE:

1. Attach stimulater device or lead wire to the garment.

2. Lay the garment flat with the four conductive silver snap zones facing up.

3. Remove liner from the GRID side of StimGel Pads.

4. Press the GRID side of the StimGel Pads over the silver snap zones on the garment.

WARNING: *The StimGel Pad must cover each conductive zone before use.*

5. Remove the remaining liner and **SAVE** for re-application to StimGel Pads in between uses.

6. Apply the garment conductive zone to your back making sure that the StimGel Pads are aligned on either side of the spine and not directly on the spine. To re-position the conductive zones, remove the garment completely. Re-adjusting with StimGel Pads attached to skin may cause Pads to roll up on themselves.

7. Turn stimulator on; adjust according to stimulator and/or clinician instructions for treatment.

8. When stimulation session is complete, turn stimulator off.

STORAGE:

9. Remove garment. The StimGel pads will lift off skin and stay attached to the garment. **StimGel Pads remain on garment for storage.**

10. Replace saved liner on gel pad. Be sure to place the “ON” side against the StimGel Pad surface.

11. Roll garment up and store in package until next treatment. Re-applying the liners, rolling up and storing the garment in re-sealable bag will prolong the life of the StimGel Pads.

ELECTRODE USE:

You can extend the use of your electrodes and StimGel Pads by following these suggestions:

- Cleanse, rinse and dry skin prior to use
- Store electrodes as indicated on packaging instructions
- If electrode gel becomes dry, adding a few drops of water to gel surface (sticky side) may help restore adhesive gel of electrode. Let the electrode sit for several minutes before applying to skin.



To reorder call Empi at **800.328.2536**

Empi StimGel Pads - 199599-001 (pack of four)